

## PRODUCT SPECIFICATION

**Spec code:** LXF-C0150  
**Product name:** Semperguard® Latex powdered  
**Date of issue:** 01/09/2014  
**Article number:** 8139 8702.  
 . =1 (size XS), 3 (size S), 5 (size M), 7 (size L), 9 (size XL)

### GENERAL INFORMATION

**Type:** single-use disposable examination or protective glove  
**Marking:** all information on dispenser box  
**Shape:** ambidextrous – straight fingers  
**Material:** Natural Rubber Latex (NRL)  
**Colour:** natural white  
**Inner:** lightly powdered  
**Outer:** no treatment  
**Cuff/Surface:** rolled cuff/fingertip textured

<b>Thickness:</b> (double measured)	<b>Finger:</b> min. 0.20 mm	<b>Palm:</b> min. 0.16 mm	<b>Cuff:</b> min. 0.11 mm
--	-----------------------------	---------------------------	---------------------------

<b>Residual powder:</b> ≤ 10 mg / glove	<b>Force at break:</b> med. 6.0 N [according to EN 455-2] (during shelf life)
---	--

<b>Length:</b> median 240 mm
------------------------------

<b>Available Sizes:</b>	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)
-------------------------	----------	---------	---------	---------	-----------

**Storage:** 5 years storage life (date of expiration see on the box). Keep area cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper-ions discolour the glove. Protect gloves against ultraviolet light sources, as sunlight and oxidizing agents. Storage above 30°C will lead to accelerated aging and should be avoided.

### STANDARDS, GUIDELINES AND QUALITY CERTIFICATION

**Quality certification:** ISO 9001, ISO 13485  
**Conformity to guidelines:** Personal Protective Equipment 89/686/EEC: CE cat. III  
 Medical Device Directive 93/42/EEC: CE Class I

**Conformity to standards:**

EN 420	EN 374-2: specified	EN 374-3: Class 3
EN 455	level 2 - AQL 1.5	Sodium Hydroxide 40%
ASTM D3578	level 3: AQL <0.65 (G1) level 2: AQL <1.5 (G1) level 1: AQL <4.0 (S4)	




**Conformity to food regulation:** 1935/2004 

### WARNING AND REPORTING SYSTEM

**Potentially allergenic ingredients:** Dithiocarbamate type, Zinc-mercaptobenzothiazol  
 (For further information, the actual version of the ingredients list is given on request.)

**Medical Device Vigilance and Reporting System:** According to the official reporting criteria of MDD, incidents caused by examination gloves or surgical gloves must be reported immediately to our medical device reporting officer. Tel.: +43 2630 310 ext. -510; Fax: -549

**Caution!** This product contains natural rubber latex which may cause allergic reactions, including anaphylactic responses!

Established (PM):	Reviewed (QA):	Approved (R&D):
 L. Millonig, MA	 DI Dr. R. Stempfer, MSc	 Dr. A. Weiner